

TECHNICAL DATA SHEET

EXTREMITY DRAPE

(EU) 2017/745 Annex XI-Part A Production Quality Assurance.

Extremity Drape Features

► Product Description	► Protective equipment used by doctors or nurses for surgical operations to prevent the possible risk of infection.
► Product Class	► (EU) 2017/745 Medical Device Regulation – Class Business Rule I
► Manufacturer's Location	► Tio Medikal 2/20 st. No.:53 (Begos 3. North Entrance, 35400 Buca OSB/Buca/İzmir)
► Purpose of usage	► The product is used for knee operations.

Quality

- Manufactured under ISO 13485:2016 and 13795-1 quality management standards.
- It has CE certificate.

Bio-Compatibility

- Does not contain latex.
- Sterilized with ethylene oxide.

Related Standard

- ISO 13485:2016 Medical Medical Devices Quality Management System
- 9001:2015 Quality Management System
- TS EN 13795 Surgical Clothing and Drapes
- TS EN ISO 11607-1:2017 Packaging for finally sterilised medical devices - Part 1: Rules for materials, sterile barrier systems and packaging systems / Products are made in accordance with the relevant standard.
- TS EN ISO 11607-2:2017 Packaging of medical devices-Finally sterilised-

Part 2: Validation requirements for forming, sealing and joining processes/
Products are made in accordance with the relevant standard.

Shelf Life

- 3 years




Figure 1. Extremity Drape Display

1. Extremity Drape Sizes




Figure 1. Extremity Drape Technical Demonstration


Available Materials: 02 Dublex(PE /Viscose or PE/Spunbond PP) , 54-56 gr/m2 double sided adhesive tape

REF . CODES	DIMENSIONS (CM)		PIECES IN BOX	
	A	B	50×80×50	40×60×40
436.0X.000.01	300	200	75	25
436.0X.001.01	320	255	75	25
436.0X.002.01	330	230	75	25
436.0X.003.01	320	220	75	25

Tolerances: +-3% (ALL SIZES ARE PRODUCED & CUSTOMISATION)

2. Extremity Pack Content

Extremity Pack	Size	QTY
Extremity drape	200*300cm	1x
Back table cover	150*200cm	1x
Op tape	10*30cm	2x
Mayo stand cover	75*145cm	1x
Skin marker and a ruler		1x
Reinforced surgical gown	L	3x
Medical towel	40*40cm	3x
non-woven wrap	100*100cm	1x



		EN 13795 requirements			
Characteristic	Test Method	High performance		Results	Main fabric
		Critical product area	Less critical product area		
Resistance to microbial penetration - Dry (CFU)	EN ISO 22612	Not required	<300	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Resistance to microbial penetration - Wet (/18)	EN ISO 22610	6,0	Not required	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Cleanliness - Microbial (CFU/100cm²)	EN ISO 11737-1	<300	<300	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Cleanliness - Particulate matter (IPM)	EN ISO 9073-10	<3,5	<3,5	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Linting (Log10 (lint count))	EN ISO 9073-10	<4,0	<4,0	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Resistance to liquid penetration	EN 13795	≥ 100	≥ 10	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Bursting strength - dry	EN 13795	≥ 40	≥ 40	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Bursting strength - wet	EN 13795	Not required	≥ 40	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Tensile strength - dry	EN 13795	≥ 20	≥ 20	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven
Tensile strength - wet	EN 13795	Not required	≥ 20	Conformed	- SS Hydrophobic Spunbond - SS Hydrophylic Spunbond - Surgigal drape and gowns-Blue/white non woven - Surgigal drape and gowns-Blue laminated non woven

BIOCOMPATIBILITY


















Biocompatibility studies has been done, and test results are given in the below table. Test results are within limits.

Table 31 Biocompatibility Study Documentation

<i>Test Name</i>	<i>Test Method</i>	<i>Test facility</i>	<i>Report No</i>	<i>Report Date</i>	<i>Result</i>
<i>Cytotoxicity</i>	ISO10993-5:2010	HACETTEPE UNIVERSITY	ARGEDS-2016/37	01.08.2016	<i>Confirmed</i>
<i>Sensitization</i>	ISO10993-10:2014	HACETTEPE UNIVERSITY	ARGEDS-2016/37	24.10.2016	<i>Confirmed</i>
<i>Skin Irritation</i>	ISO10993-10:2014	HACETTEPE UNIVERSITY	ARGEDS-2016/37	11.06.2016	<i>Confirmed</i>

Instruction for use:

1. The package of product shall be opened in sterile and aseptic conditions.
2. For a clean peel, open the package from the direction of the arrow slowly .
3. There are labels or marks on the drapes, which helps the user to follow the patient direction and unfolding directions.
4. For the incision film or adhesive tape on the drapes,first of all peel the paper carrier and fix the drape on the operational area of the patient. After then unfold the drape by following the directions.
5. Surgical drapes are ready for the operation, when they are completely unfolded.

	Do not use if package is damaged		Manufacturer		Single Use (Do not re-use)		Use by date
	Do not expose the product to sunlight.		Catalogue number		Consult instructions for use		Temperature Limitation
	Do not re-sterilize		Manufacturing Date		Sterilized using Ethyleneoxide and Single sterile barrier system		CE Marking
	Keep Dry		Batch Code		Caution		Medical device
	Unique device identifier						